

UAB GE MRI Scanners

Scanner Name	PETMR	GARDMR	LEEDSMR	TKCMR4	CVMR1
Scanner Location	Advanced Imaging Facility	Gardendale Hospital	Leeds Hospital	The Kirklin Clinic	Boshell Building
Scanner Contact	Lael Gore	Gail Reato	Kayla Freeman	Heather Edmonds	William Fisher
Phone Number	(205) 996-3272	205-996-3565	(205) 699-4180	205-801-7901	(205) 934-9906
System ID	UABMRPET	205608MR	205702450W	205801KMR3	205934CVMR
Magnetic Field Strength	3.0T	1.5T	1.5T	1.5T	1.5T
Scanner Model	PETMR	Artist	Optima 450w	Excite TwinSpeed	EchoSpeed Cvi
System Serial Number	UA0408	HM1901	HM1181	R2537	R0233
Installation Date	1/15/2015	8/15/2018	7/15/2015	3/15/2000	3/15/1999
Software Revision	MP26.0	DV26.0	DV26.0	HD16.0	LX9
Gradient Type	XRMw	XRMw	XRMw	TwinSpeed	EchoSpeed
Gradient Strength (mT/m)	44	44	44	33	33
Slew Rate (mT/m/msec)	200	200	200	80 Whole / 120 Zoom	120
Number of Receivers	32	96	32	16	4
Multi-Nuclear (MNS) Installed	N/A	N/A	N/A	N/A	N/A
Bore Diameter (cm)	60	70	70	60	60
Accreditation (MRAP)	54432-01	54520-01	53666-01	03698-04	IACMR
MRAP Expiration Date	3/8/2022	8/22/2022	5/2/2023	9/28/2024	3/31/2022
Breast MRAP (BMRAP)	N/A	51274-01	51340-01	00677-01	N/A
BMRAP Expiration Date	N/A	6/20/2022	1/16/2023	N/A	N/A