

UAB Siemens MRI Scanners

Scanner Name	HSMR1	HSMR2	TKCMR3	UABiMRI
Scanner Location	Highlands Hospital	Highlands Hospital	The Kirklin Clinic	Women & Infants Center
Scanner Contact	Dustin Johnson	Dustin Johnson	Heather Edmonds	Kayla Freeman
Phone Number	(205) 930-7355	(205) 930-7355	205-801-7901	(205)-975-3380
System ID	400-358911	400-506628	400-578948	400-562147
Magnetic Field Strength	1.5T	3.0T	3.0T	1.5T
Scanner Model	MAGNETOM Symphony	MAGNETOM Prisma	MAGNETOM Skyra	MAGNETOM Aera
System Serial Number	37265	66069	145544	142415
Installation Date	11.2009	10.2015	3.2018	2.2021
Software Revision	VB19B	VE11C	VE11C	VE11E
Gradient Type	Quantum	XR	XQ	XQ
Gradient Strength (mT/m)	45	80	30	45
Slew Rate (mT/m/msec)	200	200	125	200
Number of Receivers	18	64	64	204
Multi-Nuclear (MNS) Installed	N/A	N/A	N/A	N/A
Bore Diameter (cm)	60	60	70	70
Accreditation (MRAP)	01025-02	01025-03	03698-05	N/A
MRAP Expiration Date	9/21/2023	9/21/2023	9/28/2024	N/A
Breast MRAP (BMRAP)	N/A	N/A	N/A	N/A
BMRAP Expiration Date	N/A	N/A	N/A	N/A